



DEPARTMENT OF PHARMACEUTICS

QUOTATION ENQUIRY

Ref: IIT (BHU)/BM/IP/15-16/QTN/25

Due Date:08/02/2016
(Last date of quotation receiving)Date:18/01/2016
(Date of tender enquiry)

To,

Dear Sir,

Please submit your lowest rate for supplying the under mentioned items. Quotation in duplicate must reach us before the date marked above and should contain the following information:

1. Full specification and make of the item offered and its rate F.O.R. Varanasi/CIF New Delhi.
2. Sales tax at concessional rate as applicable to educational institution.
3. Your VAT / CST registration number, PAN & TIN numbers.
4. Conditions of supply and terms of payment.
5. If you are a manufacture of the item or if you have proprietary right over it, please mention it in the quotation and provide a certificate.
6. Please mention your agency commission in Indian Rs., if applicable (in case of imported items).
7. Please give undertaking as per annexure-I-B

Quotation must be sent in a **sealed envelope** with word "QUOTATION", our reference number, and due date as given above, clearly marked over it.

SL. NO.	Name of Items	Description / Technical Specification	Quantity
1.	Binary Gradient HPLC system with PDA detector	It should have following components: <ul style="list-style-type: none"> • Solvent delivery unit (Binary Pump): <ul style="list-style-type: none"> ❖ Flow Range: 0.0001ml/min –10.0ml/min. ❖ Flow precision: 0.06 % RSD ❖ Flow accuracy: ±1 % ❖ Max. Operating pressure: 5000 psi ❖ Plunger Capacity -10 µl • Gradient Mixer • Photodiode Array Detector <ul style="list-style-type: none"> ❖ Wavelength Range : 190-800nm ❖ No. of Diode elements : 512 ❖ Light Source: Deuterium (D2) Lamp and Tungsten (W) Lamp ❖ Wavelength accuracy: ±1 nm max. ❖ Wavelength precision: ±0.1nm max. ❖ Linearity : 2.0 AU ❖ Noise Level : 0.6 x10⁻⁵ AU 	1

		<ul style="list-style-type: none"> ❖ Wavelength accuracy: ± 1 nm max. ❖ Wavelength precision: ± 0.1 nm max. ❖ Linearity : 2.0 AU ❖ Noise Level : 0.6×10^{-5} AU <ul style="list-style-type: none"> • Manual Rheodyne Injector <ul style="list-style-type: none"> ❖ Compatible with the operating pressure of binary pump ❖ Sample loop with defined volume capacity • Syringe for sample injection • Column Oven: Temp. setting range: 4~80 °C • Ready to connect Software/ s • C-18 Column (2) • Branded personal computer & related accessories : <ul style="list-style-type: none"> ❖ Configured as i5Processor with 500GB HDD, 4GB RAM, Windows 8, LED Monitor ❖ Laser Jet Printer ❖ UPS 30 min back up (2KVA). • Warranty: <ul style="list-style-type: none"> ❖ 2 years ❖ Vendor should have service center/office in North territory of India ❖ Vendor should provide onsite training. 	
--	--	--	--

N.B. : Other terms & conditions pertaining to item mentioned above, shall be mentioned below:

The Sealed quotations will be opened on 15/02/2016 at 03:00 PM in the office of **Department of Pharmaceutics**. (This is mandatory in case of e- publishing).




HOD/COS/PI

Dr. B. Mishra, PI
Professor of Pharmaceutics
Department of Pharmaceutics
Indian Institute of Technology
(Banaras Hindu University)
Varanasi-221005