



Dr. Jeyakumar Kandasamy
Assistant Professor

DEPARTMENT OF CHEMISTRY
INDIAN INSTITUTE OF TECHNOLOGY
(BANARAS HINDU UNIVERSITY)
VARANASI-221005
UTTAR PRADESH, INDIA
Cell: +91-9794516773
Email: jeyakumar.chy@iitbhu.ac.in

IT(BHU)/CHY/2018-19/JK/GC-MS-corrigendum-2

Dated: 12.10.2018

Corrigendum

Corrigendum for tender Id: IIT(BHU)/CHY/2018-19/JK/GC-MS/110, dated: 27-09-2018 related to the purchase of Gas chromatography–Mass spectrometry.

A: Modification in Specifications

S. No.	Tender Terms/ Specifications	Modified Tender Terms/ Specifications
1.		
	Page No 20. S.No.A-6: Gas Sampling Valve for gas analysis	Optional: Gas Sampling Valve for gas analysis
2.	Page No 20. S.No.-C: Auto Sampler with 15 or more vials (1.5 to 2.0 ml) capacity	Auto Sampler with 12 or more vials (1.5 to 2.0 ml) capacity
3.	Page No 20. S.No.-D-8: Pumping System: single Turbo-molecular pump 250L/sec or more.	Pumping System: Single Turbo-molecular pump with 250L/sec or more is desirable

B: Accessories

10 KVA UPS with 60 mins backup for GC-MS may be quoted as Optional

C: Bid Submission End Date

Bid Submission End Date is extended till 01-11-2018 (06:00 PM)

C: Bid Opening Date

Bid Opening Date is 03-11-2018 (03:00 PM)

Dr. Jeyakumar Kandasamy (PI)
Dr. JEYAKUMAR .K
Assistant Professor
Department of Chemistry
Indian Institute of Technology
(Banaras Hindu University)
Varanasi - 221 005
Uttar Pradesh, India.