



भारतीय
प्रौद्योगिकी
संस्थान

IIT

INDIAN
INSTITUTE OF
TECHNOLOGY
VARANASI HINDU UNIVERSITY

Dr. Akansha Dwivedi

PhD (Pennsylvania State University)

Assistant Professor

Department of Ceramic Engineering

Indian Institute of Technology – BHU.

Varanasi – 221005, India

Email: akanshadwivedi.cer@itbhu.ac.in

Phone: +91 9455815059

Enq. No.: CER/AD/ 2017/ T-01

Enquiry Dated: August 2, 2017

Closing date: August 24, 2017

Sub: Request for submission of quotation for Ferroelectric measurement unit (Corrigendum)

This is a **Corrigendum** in an open tender published on 2-8-2017 on the same portal with the above mentioned subject. In the technical specification, please read section 2 as follows:

2. P-E Hysteresis loop measurement:

- In built voltage $\pm 10V$, $\pm 10KV$ (with external amp.), Frequency up to 5 KHz
- Arbitrary Waveform Generator output
- Pulse Widths down to at least $1\mu s$ and up to 1s
- Controlled output ramp for maximum accuracy

Akansha

Dr. Akansha Dwivedi
Assistant Professor
Department of Ceramic Engineering
Indian Institute of Technology, BHU,
Varanasi, U.P. 221005 INDIA

Forwarded

24/8/2017

HEAD/विभागाध्यक्ष
Department of Ceramic Engineering
सिरामिक अभियान्त्रिकी विभाग
Indian Institute of Technology (B.H.U.)
भारतीय प्रौद्योगिकी संस्थान (का०हि०वि०वि०)
Varanasi-221005, भारत-221005